

INSTRUCTIONS FOR USE

DEVICE DESCRIPTION

The Prodigy™ Thrombectomy System is comprised of several devices:

- Prodigy Catheter (provided accessory: Rotating Hemostasis Valve [RHV])
- Prodigy Twist™
- Prodigy Hotshot™ Controller (provided accessories: stopcock & 6" extension tubing)
- TRUVIC Generator
- TRUVIC Canister
- TRUVIC Tubeset

The Prodigy Thrombectomy System is designed to remove thrombus from the peripheral vasculature using continuous aspiration. The Prodigy Catheter targets aspiration from the TRUVIC Generator and through the Prodigy Hotshot Controller and Prodigy Catheter directly to the thrombus. The Prodigy Twist may be used to facilitate thrombus removal through the Prodigy Catheters.

The Prodigy Catheter is introduced through a guide catheter or introducer sheath and into the peripheral vasculature and guided over a guidewire to the site of the target thrombus. The Prodigy Catheter is used with the TRUVIC Generator, TRUVIC Canister, TRUVIC Tubeset and Prodigy Hotshot Controller, to aspirate thrombus. As needed, a Prodigy Twist may be introduced through the Prodigy Catheter and rotated while advancing and retracting through the Prodigy Catheter tip to assist with thrombus removal or to facilitate clearing of the thrombus from the Prodigy Catheter. The Prodigy Catheter is provided with a rotating hemostasis valve (RHV). The Prodigy Catheters and Prodigy Twists have distal radiopaque markers which are visible under fluoroscopy. The Prodigy Catheter radiopaque marker is approximately 1 mm from the end of the atraumatic catheter tip. The Prodigy Twist has two radiopaque markers contained within the polymer tip.

INDICATION FOR USE

The Prodigy Thrombectomy System is intended for the removal of fresh, soft emboli and thrombi from vessels of the peripheral arterial and venous systems. Not for use in the coronaries, pulmonary vasculature, or the neurovasculature.

CONTRAINDICATIONS

There are no known contraindications.

WARNINGS

The Prodigy Thrombectomy System should only be used by physicians who have received appropriate training in interventional techniques.

Do not advance, retract, or use any component of the Prodigy Thrombectomy System against resistance without careful assessment of the cause using fluoroscopy. If the cause cannot be determined, withdraw the device or system as a unit. Torquing, forced insertion or withdrawal of the Prodigy Catheter or Prodigy Twist against resistance may result in damage to the device or vessel; do not rotate the devices against resistance more than 1 revolution.

Do not retract the Prodigy Twist through the RHV unless the RHV is opened sufficiently to allow passage.

Verify aspiration pump is appropriate before use.

The Prodigy Catheters have been verified for use with the TRUVIC Tubeset and TRUVIC Generator. The TRUVIC Generator is capable of delivering vacuum pressures between -20 inHg and -29.9 inHg during use and is characterized by the pressure-flow performance curve presented in Figure 1. If using a vacuum pump other than the TRUVIC Generator, carefully review the vacuum pump performance parameters to ensure it can achieve the same operating vacuum pressures between -20 inHg and -29.9 inHg and corresponds to the same flow rate ranges (Figure 1).




FIGURE 1 – PUMP PRESSURE-FLOW PERFORMANCE CURVE FOR THE TRUVIC GENERATOR

PRECAUTIONS

- The device is intended for single use only. Do not resterilize or reuse. Resterilization and/or reuse may result in ineffective catheter coating lubrication, which may result in high friction and the inability to access the target vasculature location.
- Do not use kinked or damaged devices. Do not use open or damaged packages. Return all damaged devices and packaging to the manufacturer/ distributor.
- Use prior to the “Use By” date.
- Use the Prodigy Thrombectomy System in conjunction with fluoroscopic visualization.
- In order to minimize blood loss, ensure that the Prodigy Hotshot Controller vacuum switch is in the “ON” position for the minimum time needed to remove thrombus.
- The Prodigy Twist is not intended for use as a guidewire. If repositioning of the Prodigy Catheter is necessary during the revascularization procedure, such repositioning should be performed over an appropriate guidewire using standard techniques.
- Do not use automated high-pressure contrast injection equipment with the Prodigy Catheter.

POTENTIAL ADVERSE EVENTS

Possible complications include, but are not limited to, the following:

- | | |
|---|--------------------------------|
| • Acute occlusion | • Ischemia |
| • Occlusion of target artery | • Acute myocardial infarction |
| • Arrhythmia | • Infarction/Necrosis |
| • Death | • Amputation of an extremity |
| • Distal embolization | • Vessel spasm |
| • Air embolism | • Arterial injury |
| • Hematoma or hemorrhage at puncture site | • Thrombosis |
| • Peripheral vascular hemorrhage | • Renal insufficiency/Failure |
| • Hypotension | • Suboptimal revascularization |
| • Hypertension | • Device malfunction |
| • Infection, sepsis | • Local reaction |
| • Fever | |

PROCEDURE

1. Refer to Warnings, Precautions, and Potential Adverse Events prior to use.
2. As each device of the Prodigy Thrombectomy System is used, remove the device from the packaging, and inspect for damage or for kinks.
3. Prepare Prodigy Thrombectomy System devices for use by flushing with heparinized saline.
4. Prepare a guide catheter or introducer sheath according to the manufacturer's Instructions for Use.
5. Using conventional catheterization techniques under fluoroscopic guidance, place the guide catheter or introducer sheath over an appropriate guidewire and into the artery or vein to approach the thrombus site.

PRODIGY SYSTEM PREPARATION AND USE

1. Confirm vessel diameter and select an appropriately sized Prodigy Catheter and corresponding Prodigy Twist.
2. Hydrate the outer surface of the Prodigy Catheter.
3. Attach the provided RHV to the Prodigy Catheter.
4. Attach the TRUVIC Tubeset to the TRUVIC Canister and the luer connector labelled "VACUUM" of the Prodigy Hotshot Controller as indicated in Figure 2.
5. As shown in Figure 2:
 - a. Attach the stopcock provided with the Prodigy Hotshot Controller to the end of the Prodigy Hotshot Controller.
 - b. Attach the 6" extension tubing to the stopcock.
 - c. Connect the other end of the 6" extension tubing to the side port of the RHV.




FIGURE 2. ASSEMBLED PRODIGY THROMBECTOMY SYSTEM

6. Switch the vacuum switch on the Prodigy Hotshot Controller to the "OFF" position.
7. Insert the Prodigy Catheter over the guidewire and into the guide catheter, introducer sheath, or a primary Prodigy Catheter already in place (telescoping).
 - a. Use a sterile gauze pad to help grasp the distal end of the catheter while inserting through the valve.
 - b. If telescoping catheters are used, do not over tighten RHV of the primary catheter onto telescoping catheter.
8. Advance the Prodigy Catheter into the target vessel over the guidewire. Position the Prodigy Catheter tip at the thrombus site.
9. Remove the guidewire from the Prodigy Catheter.
10. Turn on the TRUVIC Generator. Confirm that the vacuum gauge reads -20 inHg or greater vacuum (Refer to the TRUVIC Generator IFU).

11. Ensure the clot container is rotated to the locked position.
12. Insert the Prodigy Twist through the RHV into the Prodigy Catheter, leaving the Prodigy Twist torquing handle at least 8 cm (3 inches) from the RHV.
13. Advance the Prodigy Catheter distal tip to be approximately 2-3 mm from the thrombus.
14. To begin aspiration thrombectomy, ensure the stopcock is open between the Prodigy Catheter and Prodigy Hotshot Controller, toggle the vacuum switch on the Prodigy Hotshot Controller to the “ON” position.
15. To assist with aspiration and removal of the thrombus, continuously rotate the Prodigy Twist in the clockwise direction (as indicated by the arrow on the torquing handle) while advancing and retracting the tip of the Prodigy Twist in and out of the Prodigy Catheter tip. As the Prodigy Twist is advanced and retracted, ensure via fluoroscopy that both radiopaque markers on the Prodigy Twist extend at least 1 cm distal to the Prodigy Catheter distal tip marker and retract proximal to the Prodigy Catheter distal tip marker.




FIGURE 3. THROMBUS ASPIRATION USING THE PRODIGY TWIST

16. To stop aspiration, toggle the vacuum switch on the Prodigy Hotshot Controller to the “OFF” position.
17. Remove the Prodigy Twist if appropriate.
18. To remove any thrombus that may remain in the Prodigy catheter, momentarily turn on aspiration by switching the vacuum switch to the ON position and ensure there is unrestricted blood flow into the clot container of the Prodigy Hotshot Controller.
19. To visualize the thrombus in the clot container of the Prodigy Hotshot Controller press the vent button to evacuate blood from the clot container.
20. If desired, thrombus in the Prodigy Hotshot Controller can be removed by pressing and holding the vent button and rotating the clot container to the unlocked position.
21. Remove the Prodigy Catheter.
22. Turn off the TRUVIC Generator.
23. If removal of additional thrombus is desired, repeat all steps above as appropriate.

STORAGE

Store in a cool, dry place.

SYMBOL GLOSSARY

Symbol	Definition
	Manufacturer
	Attention, see instructions for use

Symbol	Definition
	Sterilized by ethylene oxide
	Non pyrogenic
R_x Only	Prescription Only: Federal law restricts this device to sale by or on the order of a physician
	Do not use if package is damaged
	Do not resterilize
	Do not reuse
	Model Number
	Lot number
	Use by date

©2025 Imperative Care, Inc. IMPERATIVE CARE is a registered trademark and PRODIGY, HOTSHOT, and TRUVIC are trademarks of Imperative Care, Inc.