

Product Specifications

Zoom™ Stroke System

Customer Service Information
CustomerService@ImperativeCare.com
+1 (408) 502-7548

Catheters	Model Number	Length (CM)	MAX OD (IN / MM / F)	Distal OD (IN / MM / F)	Proximal ID (IN)	Distal ID, Actual (IN)	Effective Tip ID (IN)
Zoom System Catheters							
Zoom 88 Support	ICTC088100S	100	0.110 / 2.8 / 8	0.107 / 2.7 / 8	0.088	0.088	0.095
Zoom 88	ICTC088110	110	0.110 / 2.8 / 8	0.107 / 2.7 / 8	0.088	0.088	0.095
Zoom 7X	ICRC07X137	137	0.086 / 2.2 / 6	0.083 / 2.1 / 6	0.071	0.071	0.076
Zoom 71	ICRC071137	137	0.086 / 2.2 / 6	0.083 / 2.1 / 6	0.071	0.071	0.076
Zoom 55	ICRC055137	137	0.083 / 2.1 / 6	0.069 / 1.8 / 5	0.067	0.055	0.059
Zoom 45	ICRC045144	144	0.083 / 2.1 / 6	0.060 / 1.5 / 5	0.064	0.045	0.048
Zoom 4S	ICRC04S125	125	0.062 / 1.6 / 5	0.056 / 1.4 / 5	0.045	0.045	0.048
Zoom 4S	ICRC04S144	144	0.062 / 1.6 / 5	0.056 / 1.4 / 5	0.045	0.045	0.048
Zoom 4S	ICRC04S160	160	0.062 / 1.6 / 5	0.056 / 1.4 / 5	0.045	0.045	0.048
Zoom 35	ICRC035158	160	0.063 / 1.6 / 5	0.051 / 1.3 / 4	0.047	0.035	0.038
Radial Access							
Zoom RDL	ICTC088103R	103	0.110 / 2.8 / 8	0.107 / 2.7 / 8	0.088	0.088	
6F Insert Catheters							
Zoom SIM	IC6FSIM143	143	0.085 / 2.2 / 6	0.061 / 1.6 / 5	0.041		
Zoom VTK	IC6FVTK140	140	0.085 / 2.2 / 6	0.061 / 1.6 / 5	0.041		
Zoom VRT	IC6FVRT137	137	0.085 / 2.2 / 6	0.061 / 1.6 / 5	0.041		
Pump & Accessories		Model Number					
Zoom Aspiration Pump		TAP1B					
Zoom POD Aspiration Tubing		AFT117B					
Zoom DuoPort Canister		TDC001B					
Guidewires	Model Number	OD (IN / MM)	Length (CM)	Radiopaque Length (CM)	Hydrophilic Coating Length (CM)		
Ikon™ 24 Pro	ICGW024200P	0.024 / 0.61	200	8.5	60		
Ikon 24	ICGW024200S	0.024 / 0.61	200	8.5	60		

Important Safety Information (ISI)

Rx only.

The Zoom System consists of the following devices: (7X, 71, 55, 45, 4S, 35) Catheters, Zoom 88 Large Distal Platform (Zoom 88), Zoom 88 Large Distal Platform Support (Zoom 88 Support), TracStar LDP Large Distal Platform (TracStar LDP), Zoom Aspiration Pump or Zoom POD Aspiration Pump when used with the Zoom Aspiration Pump (or equivalent vacuum pump) with Zoom Canister or Zoom DuoPort Canister. The Zoom System, when used with the Zoom Aspiration Pump (or equivalent vacuum pump), is indicated for use in the revascularization of patients with acute ischemic stroke secondary to intracranial large vessel occlusive disease (within the internal carotid, middle cerebral – M1 and M2 segments, basilar, and vertebral arteries) within 8 hours of last known well. Patients who are ineligible for intravenous thrombolytic drug therapy or who have not responded to thrombolytic drug therapy are candidates for treatment.

The Zoom 88, Zoom 88 Support, Zoom RDL Radial Access System (Zoom RDL) are indicated for the introduction of interventional devices into the peripheral, coronary, and neuro vasculature.

The Zoom Aspiration Pump is intended for general suction use in hospitals and clinics.

The Zoom Aspiration Pump and the Zoom POD Aspiration Pump are intended to connect to the Zoom (7X, 71, 55, 45, 4S, 35) Catheter, Zoom 88, Zoom 88 Support, TracStar LDP to the Zoom Canister of the Zoom Aspiration Pump (or equivalent vacuum pump) and to allow the user to control the fluid flow.

The Zoom Canister and Zoom DuoPort Canister are intended to collect aspirated fluids for disposal and prevent fluid ingress from damaging the Zoom Aspiration Pump.

The Ikon Guidewire is indicated for general intravascular use within the peripheral and neuro vasculature to introduce and position catheters and other interventional devices. The guidewire is not intended for use in the coronary vasculature.

There are no known contraindications.

The Zoom (7X, 71, 55, 45, 4S, 35) Catheters, Zoom 88, Zoom 88 Support, TracStar LDP should only be used by physicians who have received appropriate training in interventional techniques and treatment of acute ischemic stroke. The Zoom 88, Zoom 88 Support, TracStar LDP and Zoom RDL should only be used by physicians who have received appropriate training in interventional techniques and are proficient in using guide sheaths and catheters.

The Ikon Guidewire should be used only by physicians trained in percutaneous interventional techniques and procedures.

The pump should only be used by healthcare professionals. Only physicians familiar with standard interventional techniques for aspiration catheters and aspiration pumps should use the Zoom Aspiration Pump.

For complete product information, including indications, contraindications, warnings, precautions and adverse events, see product IFU included in product packaging, contact Customer Service at +1-408-502-7548, CustomerService@ImperativeCare.com or visit bit.ly/3yWkfEJ.